

3430 Schmon Pkwy, Thorold, ON, Canada L2V 4Y6 Phone: 1-866-667-4362 • 905-227-8848

Fax: 905-227-1061 Email: techsupport@norgenbiotek.com

2X One-Step RT-PCR Master Mix (5 x 100 Reactions)

Product Insert

Product # 28115

Description:

Norgen's 2X One-Step RT-PCR Master Mix is a ready-to-use solution that contains components required for RT-PCR amplification of RNA templates. The mix includes M-MuLV reverse transcriptase, Taq DNA polymerase, dNTPs, reaction buffer, MgCl₂, KCl, and a PCR enhancer/stabilizer. The user needs only to add template, the primer set and water to the Master Mix to set up the RT-PCR reaction.

Taq Source:

An E. coli strain with a cloned Taq DNA Polymerase gene from Thermus aquaticus YT-1.

RT Source:

An E. coli strain with a cloned reverse transcriptase gene from M-MuLV

Specifications:

- · Convenience and time savings
- Cost efficient
- High sensitivity and specificity
- · Good reproducibility

Applications:

- Routine and direct RT-PCR amplification of RNA templates
- Multiple band detection or genotyping

Reagents supplied:

2X RT-PCR Master Mix (5 Vials, 100 Reactions Each). Sufficient reagent for 500 x 20 μL reactions

Storage Conditions:

2X RT.PCR Master Mix should be stored at -20°C. For everyday use an aliquot can be stored at 4°C for up to 3 months. Repeated freeze-thaw cycles are not recommended. When stored at the proper temperature this reagent is stable for at least 1 year.

Precautions and Disclaimers:

This product is designed for research purposes only. It is not intended for human or diagnostic use.

Tips for Performing PCR Reactions:

Reverse Transcriptase-Polymerase Chain Reaction (RT-PCR) is a powerful method used to amplify specific RNA transcripts using multiple cycles of a two-part process: Part I, Reverse transcription and Part II, PCR cycle containing denaturation, annealing, and extension steps. Successful RT-PCR relies on various factors, and it is important to keep a number of points in mind when performing RT-PCR:

- Using high quality, purified RNA templates greatly enhances the success of RT-PCR
- Clean, disposable gloves should be worn at all times when handling reagents, samples, pipettes, disposable tubes, etc. It is recommended that gloves are changed frequently to avoid contamination.
- 3. There should be designated RNase-free solutions, tips, tubes, pipettes, etc. for RT-PCR only.

- 4. Primers should have a melting temperature of approximately 60°C. Use primer design software.
- 5. Design primers to generate an amplicon up to 1.5 kb
- To avoid amplification of the genomic copy of the target, either treat the RNA template with DNase I or design the primer set to be flanking an intron.

Procedure

Reaction Setup Table

RT-PCR Reaction Mixture	20 μL Rxn	50 μL Rxn
2X One-Step RT-PCR Master Mix	10 μL	25 μL
Template RNA (300 fg - 300 ng)	1-2 µL	1-2 µL
Primer F (2 - 10 μM)	1 µL	1 μL
Primer R (2 - 10 μM)	1 µL	1 μL
Nuclease-Free Water	Up to 20 μL	Up to 50 μL

- Dispense either 10 μL or 25 μL of 2X One-Step RT-PCR Master Mix into the PCR tube according to the desired final volume of the RT-PCR reaction (see Reaction Setup Table above).
- Add template RNA (300 fg to 300 ng of total RNA, poly A RNA, or in vitro transcribed RNA) and both forward and reverse primers (enough for a final concentration of 100-500 nM) to the PCR tube as shown in the Reaction Setup Table.
- 3. Add nuclease-free water to bring the total volume to either 20 μL or 50 μL.
- 4. Mix the RT-PCR mixture thoroughly and spin down briefly.
- Place the PCR tubes into the PCR machine and carry out the RT-PCR according to the suggested RT-Cvcle Conditions in the tables below

Suggested PCR Cycle Conditions

PCR Cycle Step	Temperature	Time	No. of Cycles
Reverse Transcription	50°C	30 min	1
Initial Denaturation	94-95°C	5 min	1
Denaturation	94-95°C	15-30 sec	
Annealing	55-60°C	15-30 sec	30-40
Extension	65-72°C	1 minute per kb	
Final Extension	72°C	5 min	1
Hold	4-10°C	Indefinitely	1

Technical Support

Contact our Technical Support Team between the hours of 8:30 and 5:30 (Eastern Standard Time) at (905) 227-8848 or Toll Free at 1-866-667-4362. Technical support can also be obtained from our website (www.norgenbiotek.com) or through email at techsupport@norgenbiotek.com.

ISO 13485:2003, ISO 9001:2008, and ISO 15189:2012 Registered

Product of CANADA